
Midmark M11 - v1.0.0
Total Cycles: 4995

Name: RITTER 1
Sterilizer ID: V2178198
10/15/2024 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.4 F 0.0
2:00 175.1 F 0.2
4:00 177.8 F 1.1
6:00 198.6 F 5.3
8:00 214.2 F 9.3
10:00 225.3 F 11.4
12:00 230.2 F 8.6
14:00 241.6 F 12.9
16:00 240.3 F 11.2
18:00 238.1 F 9.5
20:00 251.1 F 15.8
22:00 261.9 F 22.0
24:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 28.9
2:00 271.6 F 28.2
3:00 272.5 F 29.0
4:00 272.4 F 28.7

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.3 F 0.0
5:00 211.2 F 0.0
10:00 229.1 F 0.0
15:00 229.3 F 0.0
20:00 229.6 F 0.0
25:00 225.2 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:34
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:06
CYCLE COMPLETE

10/15/2024 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____