
Midmark M11 - v1.0.0
Total Cycles: 2040

Name: RITTER 1
Sterilizer ID: V2178198
06/21/2022 10:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.8 F 0.0
2:00 160.7 F 1.8
4:00 177.4 F 5.7
6:00 198.8 F 10.7
8:00 219.3 F 4.1
10:00 231.2 F 9.3
12:00 238.6 F 10.6
14:00 240.3 F 11.1
16:00 236.8 F 9.0
18:00 250.7 F 15.5
20:00 262.5 F 22.3
22:00 272.5 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.2
1:00 272.6 F 29.0
2:00 271.6 F 28.2
3:00 271.9 F 28.9
4:00 273.2 F 29.3

Min 270.9 F 28.1
Max 273.2 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 210.3 F 0.0
10:00 197.4 F 0.0
15:00 199.6 F 0.0
20:00 210.2 F 0.0
25:00 247.3 F 0.0
30:00 220.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:36
CYCLE COMPLETE

06/21/2022 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____