
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003523
9:12 26/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:04
STERILIZING 4:23
136.4 C 223kPa 4:23
136.2 C 223kPa 4:53
136.1 C 222kPa 5:23
136.4 C 224kPa 5:53
136.4 C 222kPa 6:23
136.1 C 220kPa 6:53
136.1 C 222kPa 7:23
136.1 C 220kPa 7:53
136.1 C 220kPa 7:53
VENTING 7:53
AIR DRYING 8:45
CYCLE COMPLETE 8:48

Digital Signature #
10A4D4EA4C8EFB41
