

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1397

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2021 08:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.6 F 0.0  
2:00 71.9 F 0.0  
4:00 79.8 F 0.0  
6:00 117.7 F 0.0  
8:00 161.7 F 0.1  
10:00 183.2 F 1.2  
12:00 197.6 F 4.6  
14:00 209.7 F 7.7  
16:00 220.9 F 11.2  
18:00 223.0 F 5.9  
20:00 232.7 F 9.0  
22:00 242.4 F 13.2  
24:00 239.1 F 10.1  
26:00 237.2 F 8.5  
28:00 249.0 F 14.5  
30:00 261.3 F 21.4  
32:00 272.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.1  
1:00 272.4 F 28.4  
2:00 272.3 F 28.6  
3:00 272.1 F 28.2  
4:00 272.7 F 28.8  
  
Min 271.6 F 27.9  
Max 272.9 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 159.7 F 0.0  
5:00 210.4 F 0.0  
10:00 212.7 F 0.0  
15:00 193.2 F 0.0  
20:00 191.2 F 0.0  
25:00 179.0 F 0.0  
30:00 174.0 F 0.0  
  
Drying Complete

FILLING: 1:02  
HEATING: 32:25  
STERILIZING: 4:00  
VENTING: 2:58  
DRYING: 30:00  
TOTAL CYCLE: 01:10:26  
CYCLE COMPLETE

11/29/2021 09:15 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_