

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000982  
15:31 14/12/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:33  
PRESSURIZING 3:34  
STERILIZING 4:32  
136.4 C 223kPa 4:32  
136.3 C 225kPa 5:03  
136.4 C 224kPa 5:33  
136.3 C 223kPa 6:03  
136.3 C 224kPa 6:33  
136.1 C 219kPa 7:03  
136.2 C 222kPa 7:33  
136.2 C 222kPa 8:03  
136.2 C 222kPa 8:03  
VENTING 8:03  
AIR DRYING 8:57  
CYCLE COMPLETE 8:58

-----  
Digital Signature #  
3671505AB5E3021F  
-----