

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000959  
9:31 07/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:07  
PRESSURIZING 4:09  
STERILIZING 5:34  
136.0 C 220kPa 5:34  
136.4 C 232kPa 6:04  
136.3 C 232kPa 6:34  
136.2 C 229kPa 7:04  
136.4 C 232kPa 7:34  
136.4 C 232kPa 8:04  
136.4 C 230kPa 8:34  
136.4 C 232kPa 9:04  
136.4 C 232kPa 9:04  
VENTING 9:04  
AIR DRYING 10:06  
CYCLE COMPLETE 25:06  
-----

Digital Signature #  
4C0BFE80D9CA78A1  
-----