
Midmark M11 - v1.0.5
Total Cycles: 3495

Name: RITTER2
Sterilizer ID: V2313228
04/03/2024 03:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.4 F 0.0
2:00 105.7 F 0.0
4:00 147.2 F 0.2
6:00 169.8 F 2.6
8:00 179.6 F 6.3
10:00 196.6 F 10.4
12:00 210.6 F 5.2
14:00 223.3 F 7.8
16:00 234.8 F 12.3
18:00 235.1 F 9.5
20:00 238.3 F 9.7
22:00 247.4 F 14.4
24:00 259.4 F 21.0
26:00 270.4 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.4
2:00 271.6 F 29.0
3:00 272.8 F 29.5
4:00 271.6 F 28.8

Min 271.5 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 200.0 F 0.0
10:00 195.8 F 0.0
15:00 197.6 F 0.0
20:00 195.4 F 0.0
25:00 193.1 F 0.0
30:00 190.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:50
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:05:01
CYCLE COMPLETE

04/03/2024 05:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____