
Midmark M11 - v1.0.0
Total Cycles: 2683

Name: RITTER 1
Sterilizer ID: V2178198
12/15/2022 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 182.0 F 3.3
4:00 191.3 F 8.0
6:00 211.1 F 13.3
8:00 224.9 F 8.6
10:00 237.9 F 11.2
12:00 240.1 F 11.5
14:00 238.5 F 9.8
16:00 252.5 F 16.9
18:00 264.6 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.6 F 29.5
2:00 272.5 F 29.2
3:00 271.8 F 28.7
4:00 271.3 F 28.6

Min 271.3 F 28.5
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 207.7 F 0.0
10:00 221.6 F 0.0
15:00 224.3 F 0.0
20:00 225.4 F 0.0
25:00 229.6 F 0.0
30:00 233.0 F 0.0

Drying Complete

FILLING: 1:46
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

12/15/2022 12:49 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____