
Midmark M11 - v1.0.0
Total Cycles: 1396

Name: RITTER 1
Sterilizer ID: V2178198
11/24/2021 04:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.8 F 0.0
2:00 194.4 F 3.2
4:00 192.9 F 7.1
6:00 210.4 F 11.8
8:00 219.8 F 6.1
10:00 232.2 F 11.1
12:00 232.3 F 7.7
14:00 245.0 F 13.1
16:00 243.3 F 11.9
18:00 255.1 F 18.0
20:00 265.5 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.4 F 29.2
2:00 272.6 F 28.9
3:00 271.8 F 28.4
4:00 271.5 F 28.6

Min 271.3 F 28.3
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 222.2 F 0.0
10:00 224.9 F 0.0
15:00 220.5 F 0.0
20:00 220.6 F 0.0
25:00 222.4 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:43
CYCLE COMPLETE

11/24/2021 05:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____