
Midmark M11 - v1.0.0
Total Cycles: 1317

Name: RITTER 1
Sterilizer ID: V2178198
10/28/2021 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 178.8 F 3.0
4:00 200.1 F 7.4
6:00 217.8 F 12.6
8:00 228.5 F 7.8
10:00 242.2 F 13.6
12:00 242.3 F 11.6
14:00 240.4 F 10.4
16:00 254.5 F 17.5
18:00 266.4 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.2
1:00 272.6 F 28.8
2:00 271.8 F 28.4
3:00 272.7 F 28.8
4:00 271.6 F 28.2

Min 271.6 F 28.1
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.2 F 0.0
5:00 219.0 F 0.0
10:00 204.0 F 0.0
15:00 203.4 F 0.0
20:00 205.7 F 0.0
25:00 221.0 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 2:30
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:10
CYCLE COMPLETE

10/28/2021 03:27 PM
ACCEPTED: X
REJECTED:
Operator: U04