

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3080

Name: RITTER 1  
Sterilizer ID: V2178198  
04/10/2023 10:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.6 F 0.0  
2:00 155.1 F 2.5  
4:00 171.2 F 5.5  
6:00 197.8 F 10.4  
8:00 216.5 F 5.0  
10:00 227.4 F 9.5  
12:00 234.2 F 9.3  
14:00 241.1 F 11.6  
16:00 238.9 F 9.8  
18:00 252.4 F 16.4  
20:00 264.0 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.7  
2:00 272.1 F 28.9  
3:00 272.1 F 28.5  
4:00 272.1 F 28.9  
  
Min 271.3 F 28.2  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.6 F 0.0  
5:00 235.7 F 0.0  
10:00 230.8 F 0.0  
15:00 234.3 F 0.0  
20:00 215.3 F 0.0  
25:00 210.0 F 0.0  
30:00 208.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 21:58  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:08  
CYCLE COMPLETE

04/10/2023 11:11 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_