
Midmark M11 - v1.0.5
Total Cycles: 2570

Name: RITTER2
Sterilizer ID: V2313228
06/26/2023 03:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 207.4 F 0.0
2:00 199.7 F 3.9
4:00 202.1 F 9.1
6:00 220.9 F 8.8
8:00 234.0 F 10.3
10:00 234.2 F 8.2
12:00 241.4 F 11.6
14:00 246.0 F 13.7
16:00 258.4 F 20.4
18:00 269.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 29.3
1:00 272.4 F 30.0
2:00 272.6 F 29.7
3:00 271.5 F 29.3
4:00 272.9 F 30.1

Min 271.4 F 29.2
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 231.2 F 0.0
10:00 221.9 F 0.0
15:00 209.8 F 0.0
20:00 205.3 F 0.0
25:00 202.3 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 2:21
HEATING: 19:41
STERILIZING: 4:00
VENTING: 3:27
DRYING: 30:00
TOTAL CYCLE: 00:59:29
CYCLE COMPLETE

06/26/2023 04:22 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____