

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002556  
7:51 28/10/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:22  
PRESSURIZING 2:25  
STERILIZING 3:28  
136.5 C 223kPa 3:28  
136.3 C 223kPa 3:58  
136.4 C 223kPa 4:28  
136.4 C 219kPa 4:58  
135.9 C 221kPa 5:28  
136.3 C 223kPa 5:58  
136.2 C 222kPa 6:28  
136.4 C 221kPa 6:58  
136.4 C 221kPa 6:58  
VENTING 6:58  
AIR DRYING 7:52  
CYCLE COMPLETE 7:55

-----  
Digital Signature #  
5D15A18F743BB934  
-----