
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001611
11:42 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:37
STERILIZING 3:39
136.4 C 225kPa 3:39
136.2 C 225kPa 4:09
136.4 C 226kPa 4:39
136.4 C 221kPa 5:09
136.3 C 225kPa 5:39
136.4 C 226kPa 6:09
136.0 C 222kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 8:05
CYCLE COMPLETE 8:16

Digital Signature #
41732A975055CCF1
