
Midmark M11 - v1.0.5
Total Cycles: 3970

Name: RITTER2
Sterilizer ID: V2313228
09/24/2024 11:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.2 F 0.0
2:00 92.2 F 0.0
4:00 103.9 F 0.1
6:00 135.5 F 0.3
8:00 162.8 F 2.8
10:00 183.0 F 5.6
12:00 199.6 F 8.9
14:00 208.5 F 13.0
16:00 217.8 F 6.6
18:00 231.7 F 11.0
20:00 231.8 F 7.7
22:00 243.0 F 13.2
24:00 242.0 F 11.9
26:00 253.5 F 17.7
28:00 263.8 F 23.9
30:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.1
1:00 271.7 F 29.5
2:00 272.7 F 29.7
3:00 271.8 F 29.2
4:00 271.5 F 29.4

Min 271.4 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.7 F 0.0
5:00 191.0 F 0.0
10:00 198.4 F 0.0
15:00 193.7 F 0.0
20:00 196.7 F 0.0
25:00 196.3 F 0.0
30:00 198.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 30:08
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

09/24/2024 12:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____