
Midmark M11 - v1.0.5
Total Cycles: 4475

Name: RITTER2
Sterilizer ID: V2313228
03/31/2025 04:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.2 F 0.0
2:00 114.1 F 0.8
4:00 119.9 F 2.6
6:00 192.0 F 5.3
8:00 189.9 F 9.1
10:00 202.6 F 12.9
12:00 212.7 F 4.6
14:00 221.6 F 8.0
16:00 233.5 F 11.8
18:00 231.8 F 8.3
20:00 243.4 F 13.3
22:00 241.4 F 11.6
24:00 252.3 F 17.1
26:00 262.1 F 22.9
28:00 271.2 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.8 F 30.0
1:00 272.2 F 30.0
2:00 272.0 F 30.4
3:00 271.8 F 30.4
4:00 272.1 F 30.4

Min 271.7 F 29.9
Max 272.8 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 195.4 F 0.0
10:00 205.8 F 0.0
15:00 204.7 F 0.0
20:00 195.7 F 0.0
25:00 193.9 F 0.0
30:00 190.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 29:16
STERILIZING: 4:00
VENTING: 3:58
DRYING: 30:00
TOTAL CYCLE: 01:08:14
CYCLE COMPLETE

03/31/2025 05:40 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____