
Midmark M11 - v1.0.0
Total Cycles: 4133

Name: RITTER 1
Sterilizer ID: V2178198
12/19/2023 08:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.5 F 0.0
2:00 78.0 F 0.1
4:00 116.5 F 0.9
6:00 118.7 F 3.0
8:00 148.0 F 5.8
10:00 169.9 F 9.0
12:00 188.9 F 12.7
14:00 209.7 F 3.3
16:00 217.2 F 6.5
18:00 227.7 F 9.9
20:00 233.8 F 10.3
22:00 236.7 F 9.8
24:00 238.4 F 9.4
26:00 245.5 F 13.0
28:00 257.0 F 19.0
30:00 267.2 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.2 F 29.1
2:00 271.9 F 28.4
3:00 272.3 F 29.1
4:00 272.2 F 28.6

Min 271.1 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.9 F 0.0
5:00 238.7 F 0.0
10:00 201.7 F 0.0
15:00 228.3 F 0.0
20:00 217.0 F 0.0
25:00 201.9 F 0.0
30:00 188.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:22
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:38
CYCLE COMPLETE

12/19/2023 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____