

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3760

Name: RITTER2  
Sterilizer ID: V2313228  
07/08/2024 01:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.9 F 0.0  
2:00 113.6 F 1.0  
4:00 129.0 F 3.3

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 136F  
CHAMBER PRESS 4.1PSI  
\*\* ITEMS NOT STERILE \*\*

07/08/2024 01:18 PM