
Midmark M11 - v1.0.5
Total Cycles: 4449

Name: RITTER2
Sterilizer ID: V2313228
03/24/2025 03:26 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.7 F 0.0
2:00 125.8 F 1.3
4:00 138.6 F 4.1
6:00 168.7 F 7.8
8:00 190.4 F 12.0
10:00 219.0 F 5.1
12:00 217.8 F 9.3
14:00 230.4 F 11.3
16:00 236.4 F 11.0
18:00 235.0 F 9.1
20:00 248.8 F 15.4
22:00 259.9 F 21.7
24:00 270.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 30.5
1:00 272.6 F 30.9
2:00 272.7 F 30.8
3:00 272.0 F 30.8
4:00 272.2 F 30.8

Min 271.9 F 30.4
Max 273.1 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 239.2 F 0.0
10:00 224.4 F 0.0
15:00 230.1 F 0.0
20:00 226.2 F 0.0
25:00 222.6 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 25:39
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:04:36
CYCLE COMPLETE

03/24/2025 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____