
Midmark M11 - v1.0.5
Total Cycles: 4348

Name: RITTER2
Sterilizer ID: V2313228
02/10/2025 05:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.3 F 0.0
2:00 168.3 F 3.2
4:00 187.9 F 7.7
6:00 210.9 F 12.9
8:00 224.6 F 7.4
10:00 237.7 F 12.6
12:00 237.0 F 10.0
14:00 236.6 F 8.6
16:00 249.8 F 15.5
18:00 261.9 F 22.5
20:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.8 F 29.8
2:00 272.0 F 28.9
3:00 272.4 F 29.7
4:00 272.9 F 29.5

Min 271.6 F 28.8
Max 273.1 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.6 F 0.0
5:00 226.1 F 0.0
10:00 243.4 F 0.0
15:00 232.5 F 0.0
20:00 226.0 F 0.0
25:00 227.1 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:42
HEATING: 20:23
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

02/10/2025 06:37 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____