
Midmark M11 - v1.0.5
Total Cycles: 3931

Name: RITTER2
Sterilizer ID: V2313228
09/11/2024 02:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.6 F 0.0
2:00 147.0 F 2.0
4:00 155.9 F 4.9
6:00 186.7 F 8.8
8:00 206.5 F 13.0
10:00 216.3 F 6.9
12:00 229.3 F 11.4
14:00 231.2 F 7.7
16:00 241.2 F 13.1
18:00 241.6 F 11.6
20:00 252.9 F 17.6
22:00 263.5 F 23.9
24:00 272.2 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.1
1:00 272.2 F 29.8
2:00 271.5 F 29.2
3:00 271.4 F 29.6
4:00 272.7 F 30.5

Min 271.0 F 29.0
Max 273.0 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 225.2 F 0.0
10:00 230.2 F 0.0
15:00 239.7 F 0.0
20:00 229.5 F 0.0
25:00 221.0 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 24:15
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:03
CYCLE COMPLETE

09/11/2024 03:55 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____