
Midmark M11 - v1.0.5
Total Cycles: 1329

Name: RITTER2
Sterilizer ID: V2313228
07/21/2022 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 176.0 F 0.2
4:00 204.9 F 4.2
6:00 220.8 F 9.5
8:00 227.6 F 8.3
10:00 234.8 F 10.5
12:00 234.4 F 8.1
14:00 242.4 F 11.6
16:00 245.3 F 13.3
18:00 257.5 F 19.9
20:00 268.4 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.5
1:00 272.3 F 29.4
2:00 272.6 F 29.7
3:00 272.5 F 30.0
4:00 271.7 F 29.6

Min 271.7 F 29.4
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.1 F 0.0
5:00 219.1 F 0.0
10:00 225.8 F 0.0
15:00 228.1 F 0.0
20:00 226.0 F 0.0
25:00 226.5 F 0.0
30:00 173.6 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 21:56
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

07/21/2022 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____