
Midmark M11 - v1.0.5
Total Cycles: 2799

Name: RITTER2
Sterilizer ID: V2313228
08/25/2023 03:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 158.5 F 0.0
4:00 181.2 F 1.7
6:00 207.2 F 5.4
8:00 212.0 F 9.7
10:00 225.6 F 9.2
12:00 230.8 F 10.0
14:00 235.5 F 9.0
16:00 245.4 F 13.0
18:00 244.8 F 12.9
20:00 256.7 F 19.2
22:00 267.3 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.5 F 29.2
2:00 271.2 F 29.0
3:00 272.2 F 29.2
4:00 271.1 F 28.9

Min 271.0 F 28.6
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.2 F 0.0
5:00 238.3 F 0.0
10:00 218.9 F 0.0
15:00 186.7 F 0.0
20:00 185.5 F 0.0
25:00 175.4 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:26
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:01:54
CYCLE COMPLETE

08/25/2023 04:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____