
Midmark M11 - v1.0.0
Total Cycles: 2443

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2022 02:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 210.2 F 0.0
2:00 204.3 F 4.6
4:00 203.4 F 9.6
6:00 220.9 F 6.3
8:00 233.5 F 10.1
10:00 237.5 F 8.1
12:00 244.2 F 12.6
14:00 244.3 F 12.7
16:00 257.0 F 19.5
18:00 268.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.4 F 29.4
2:00 271.7 F 28.6
3:00 272.3 F 29.5
4:00 272.1 F 29.1

Min 271.4 F 28.6
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 228.5 F 0.0
10:00 236.7 F 0.0
15:00 234.8 F 0.0
20:00 227.1 F 0.0
25:00 230.4 F 0.0
Dry Cycle Aborted

FILLING: 2:14
HEATING: 19:14
STERILIZING: 4:00
VENTING: 3:14
DRYING: 26:39
TOTAL CYCLE: 00:55:22

10/17/2022 03:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____