
Midmark M11 - v1.0.0
Total Cycles: 4196

Name: RITTER 1
Sterilizer ID: V2178198
01/11/2024 02:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.9 F 0.0
2:00 168.4 F 3.3
4:00 182.0 F 7.1
6:00 202.1 F 12.0
8:00 218.7 F 6.2
10:00 228.9 F 11.0
12:00 230.8 F 7.2
14:00 242.8 F 13.1
16:00 241.8 F 11.5
18:00 253.5 F 17.6
20:00 264.6 F 24.0
22:00 271.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.2 F 29.4
2:00 272.5 F 29.3
3:00 271.8 F 28.9
4:00 271.1 F 28.5

Min 271.1 F 28.5
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 221.5 F 0.0
10:00 220.8 F 0.0
15:00 226.0 F 0.0
20:00 224.0 F 0.0
25:00 223.4 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 22:07
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:01:29
CYCLE COMPLETE

01/11/2024 03:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____