
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001643
16:05 30/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:39
136.0 C 220kPa 3:39
136.3 C 226kPa 4:09
136.3 C 224kPa 4:39
136.2 C 224kPa 5:09
136.3 C 224kPa 5:39
136.4 C 225kPa 6:09
136.2 C 223kPa 6:39
136.4 C 221kPa 7:09
136.4 C 221kPa 7:09
VENTING 7:09
AIR DRYING 8:01
CYCLE COMPLETE 8:22

Digital Signature #
0219F8C90E51D103
