
Midmark M11 - v1.0.0
Total Cycles: 4664

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2024 09:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 168.8 F 2.8
4:00 173.3 F 6.7
6:00 199.4 F 11.2
8:00 213.3 F 4.9
10:00 228.6 F 10.1
12:00 234.4 F 7.8
14:00 241.6 F 12.2
16:00 240.3 F 10.5
18:00 253.7 F 17.1
20:00 265.2 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.9
2:00 271.5 F 28.3
3:00 272.6 F 28.9
4:00 271.7 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.2 F 0.0
5:00 214.3 F 0.0
10:00 221.5 F 0.0
15:00 221.8 F 0.0
20:00 232.5 F 0.0
25:00 229.3 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:00:12
CYCLE COMPLETE

06/20/2024 10:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____