

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3538

Name: RITTER 1  
Sterilizer ID: V2178198  
07/17/2023 04:44 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.2 F 0.0  
2:00 176.8 F 1.3  
4:00 198.0 F 5.0  
6:00 210.8 F 9.5  
8:00 226.6 F 9.1  
10:00 230.1 F 8.9  
12:00 240.6 F 11.6  
14:00 241.7 F 11.8  
16:00 240.3 F 10.6  
18:00 253.3 F 17.0  
20:00 264.5 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.1 F 28.7  
2:00 271.8 F 28.8  
3:00 271.9 F 28.5  
4:00 272.0 F 29.0  
  
Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 227.6 F 0.0  
10:00 232.2 F 0.0  
15:00 228.3 F 0.0  
20:00 226.1 F 0.0  
25:00 225.8 F 0.0  
30:00 221.9 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 21:56  
STERILIZING: 4:00  
VENTING: 3:37  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

07/17/2023 05:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_