
Midmark M11 - v1.0.5
Total Cycles: 774

Name: RITTER2
Sterilizer ID: V2313228
02/08/2022 04:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.9 F 0.0
2:00 174.7 F 3.8
4:00 195.3 F 8.7
6:00 214.0 F 9.3
8:00 229.0 F 9.4
10:00 234.7 F 9.4
12:00 243.9 F 12.7
14:00 245.1 F 13.3
16:00 258.2 F 20.4
18:00 269.8 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.4
2:00 271.3 F 29.1
3:00 272.5 F 29.5
4:00 271.5 F 28.8

Min 271.3 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.7 F 0.0
5:00 208.8 F 0.0
10:00 218.6 F 0.0
15:00 224.3 F 0.0
20:00 226.4 F 0.0
25:00 224.9 F 0.0
30:00 219.6 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

02/08/2022 05:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____