
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001922
12:49 07/08/2024
AIR DRYING ONLY
60 min. MAXIMUM
CYCLE START 0:00
AIR DRYING 0:10
CYCLE COMPLETE
1:00:10

Digital Signature #
38149A1B1F620758
