

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002440  
9:00 23/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:12  
PRESSURIZING 3:17  
STERILIZING 4:19  
136.4 C 223kPa 4:19  
136.3 C 223kPa 4:49  
136.4 C 225kPa 5:19  
136.4 C 224kPa 5:49  
136.4 C 225kPa 6:19  
136.1 C 223kPa 6:49  
136.3 C 223kPa 7:19  
136.1 C 222kPa 7:49  
136.1 C 222kPa 7:49  
VENTING 7:49  
AIR DRYING 8:40  
CYCLE COMPLETE 38:41  
-----

Digital Signature #  
47DE7E1F9F2AC900  
-----