
Midmark M11 - v1.0.0
Total Cycles: 5236

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2025 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.4 F 0.0
2:00 162.8 F 2.5
4:00 183.9 F 6.4
6:00 201.5 F 10.7
8:00 216.3 F 4.8
10:00 229.2 F 8.8
12:00 238.2 F 11.5
14:00 239.5 F 10.9
16:00 236.4 F 9.0
18:00 250.6 F 15.5
20:00 262.0 F 22.0
22:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 271.8 F 28.4
2:00 272.2 F 28.7
3:00 272.0 F 28.9
4:00 271.4 F 28.1

Min 271.3 F 28.1
Max 272.4 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.0 F 0.0
5:00 212.1 F 0.0
10:00 207.9 F 0.0
15:00 227.8 F 0.0
20:00 232.0 F 0.0
25:00 227.9 F 0.0
30:00 229.1 F 0.0
Drying Complete

FILLING: 1:17
HEATING: 22:21
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:01:26
CYCLE COMPLETE

01/10/2025 04:30 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____