
Midmark M11 - v1.0.0
Total Cycles: 1554

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 181.5 F 4.0
4:00 195.6 F 8.7
6:00 217.4 F 7.2
8:00 229.2 F 9.4
10:00 238.1 F 9.3
12:00 242.4 F 12.6
14:00 242.1 F 11.4
16:00 255.6 F 18.4
18:00 267.4 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.6 F 29.1
2:00 271.7 F 28.4
3:00 272.6 F 29.3
4:00 272.3 F 28.8

Min 271.5 F 28.3
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 228.5 F 0.0
10:00 234.6 F 0.0
15:00 228.1 F 0.0
20:00 227.4 F 0.0
25:00 231.6 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:39
HEATING: 19:21
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

01/24/2022 12:10 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____