

-----  
Midmark M11 - v1.0.5  
Total Cycles: 638

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2021 02:48 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.2 F 0.0  
2:00 151.7 F 1.7  
4:00 179.3 F 5.4  
6:00 193.1 F 10.4  
8:00 214.8 F 4.9  
10:00 227.6 F 9.0  
12:00 237.1 F 11.3  
14:00 237.9 F 11.4  
16:00 238.1 F 9.9  
18:00 250.3 F 16.5  
20:00 261.6 F 23.0  
22:00 271.9 F 30.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.7  
1:00 272.4 F 30.5  
2:00 272.4 F 30.1  
3:00 271.4 F 29.5  
4:00 271.7 F 30.1

Min 271.3 F 29.5  
Max 272.9 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.7 F 0.0  
5:00 201.0 F 0.0  
10:00 208.2 F 0.0  
15:00 208.5 F 0.0  
20:00 209.5 F 0.0  
25:00 210.2 F 0.0  
30:00 211.5 F 0.0  
Drying Complete

FILLING: 1:47  
HEATING: 22:36  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:01:38  
CYCLE COMPLETE

12/15/2021 03:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_