

-----  
Midmark M11 - v1.0.0  
Total Cycles: 702

Name: MIDMARK  
Sterilizer ID: V2178198  
02/25/2021 10:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	114.9 F	0.0
2:00	125.1 F	0.1
4:00	134.7 F	2.1
6:00	162.2 F	5.3
8:00	191.0 F	9.8
10:00	219.2 F	5.4
12:00	220.7 F	8.3
14:00	234.3 F	13.4
16:00	237.1 F	10.5
18:00	234.1 F	8.0
20:00	249.8 F	14.9
22:00	262.1 F	21.8
24:00	272.6 F	28.5

STERILIZING

mm:ss	Degrees	PSI
0:00	271.7 F	28.4
1:00	271.8 F	28.6
2:00	272.0 F	28.8
3:00	272.1 F	28.9
4:00	272.0 F	28.9

Min 271.6 F 28.3  
Max 272.7 F 28.9

VENTING CHAMBER

mm:ss	Drying Start	Degrees	PSI
0:00		193.9 F	0.0
5:00		227.3 F	0.0
10:00		199.5 F	0.0
15:00		195.5 F	0.0
20:00		191.0 F	0.0
25:00		190.7 F	0.0
30:00		234.1 F	0.0

Drying Complete

FILLING: 0:56  
HEATING: 24:52  
STERILIZING: 4:00  
VENTING: 2:56  
DRYING: 30:00  
TOTAL CYCLE: 01:02:44  
CYCLE COMPLETE

02/25/2021 11:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_