

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1820

Name: RITTER2  
Sterilizer ID: V2313228  
12/07/2022 04:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 177.5 F 0.0  
2:00 178.3 F 4.3  
4:00 194.9 F 9.1  
6:00 214.2 F 7.2  
8:00 228.7 F 9.7  
10:00 232.7 F 8.5  
12:00 243.2 F 12.6  
14:00 244.8 F 13.1  
16:00 257.7 F 20.2  
18:00 269.6 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.7 F 29.9  
2:00 271.7 F 29.1  
3:00 272.4 F 30.1  
4:00 272.6 F 29.7  
  
Min 271.4 F 29.0  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 227.6 F 0.0  
10:00 240.5 F 0.0  
15:00 238.5 F 0.0  
20:00 226.9 F 0.0  
25:00 221.2 F 0.0  
30:00 221.3 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 18:56  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:57:41  
CYCLE COMPLETE

12/07/2022 05:08 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_