
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001457
12:59 09/29/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:19
PRESSURIZING 7:12
STERILIZING 8:50
136.0 C 220kPa 8:50
135.8 C 217kPa 9:20
136.4 C 223kPa 9:50
136.4 C 222kPa 10:20
136.4 C 222kPa 10:50
136.4 C 222kPa 11:20
136.4 C 222kPa 11:50
135.9 C 220kPa 12:20
VENTING 12:20
AIR DRYING 13:26
CYCLE COMPLETE 20:49

Digital Signature #
01CFC33732190AD2
