
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001329
13:42 31/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:01
PRESSURIZING 4:04
STERILIZING 5:19
136.0 C 220kPa 5:19
136.2 C 224kPa 5:49
136.4 C 224kPa 6:19
136.4 C 225kPa 6:49
136.4 C 223kPa 7:19
136.3 C 224kPa 7:49
136.1 C 224kPa 8:19
136.2 C 223kPa 8:49
136.2 C 223kPa 8:49
VENTING 8:49
AIR DRYING 9:42
CYCLE COMPLETE 14:37

Digital Signature #
936E3BA8A8AAA137
