

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000147  
11:07 14/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:36  
PRESSURIZING 4:36  
STERILIZING 5:48  
136.0 C 222kPa 5:48  
136.4 C 224kPa 6:18  
136.4 C 226kPa 6:48  
136.4 C 224kPa 7:18  
136.2 C 224kPa 7:48  
136.4 C 226kPa 8:18  
136.2 C 223kPa 8:48  
136.4 C 226kPa 9:18  
136.4 C 226kPa 9:18  
VENTING 9:18  
AIR DRYING 10:19  
CYCLE COMPLETE 14:55  
-----

Digital Signature #  
2217C327A1C043D2  
-----