

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001562  
9:53 11/05/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:41  
PRESSURIZING 6:19  
STERILIZING 7:50  
136.0 C 220kPa 7:50  
135.7 C 217kPa 8:21  
136.4 C 219kPa 8:51  
136.2 C 221kPa 9:21  
136.2 C 221kPa 9:51  
136.5 C 220kPa 10:21  
136.5 C 223kPa 10:51  
136.0 C 220kPa 11:20  
VENTING 11:21  
AIR DRYING 12:24  
CYCLE COMPLETE 17:04

-----  
Digital Signature #  
AF46DE6F65734A02  
-----