
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 008122
10:45 02/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:54
PRESSURIZING 2:56
STERILIZING 4:02
136.0 C 220kPa 4:02
136.0 C 222kPa 4:32
136.2 C 221kPa 5:02
136.2 C 224kPa 5:32
136.2 C 221kPa 6:02
136.2 C 219kPa 6:32
136.3 C 223kPa 7:02
136.4 C 224kPa 7:32
136.4 C 224kPa 7:32
VENTING 7:32
AIR DRYING 8:26
CYCLE COMPLETE 8:35

Digital Signature #
10348E423B91E8B2
