
Midmark M11 - v1.0.0
Total Cycles: 2364

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 04:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.5 F 0.0
2:00 157.3 F 3.0
4:00 171.5 F 6.8
6:00 220.6 F 11.1
8:00 227.1 F 5.5
10:00 237.4 F 10.5
12:00 232.5 F 7.0
14:00 244.5 F 12.9
16:00 241.8 F 11.1
18:00 254.6 F 17.6
20:00 265.8 F 24.4
22:00 272.2 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.8
1:00 272.2 F 28.9
2:00 272.7 F 29.3
3:00 271.3 F 28.9
4:00 272.4 F 28.9

Min 271.1 F 28.7
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.4 F 0.0
5:00 217.7 F 0.0
10:00 225.7 F 0.0
15:00 237.3 F 0.0
20:00 239.6 F 0.0
25:00 229.9 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

09/22/2022 05:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____