

-----  
Midmark M11 - v1.0.0  
Total Cycles: 657

Name: MIDMARK  
Sterilizer ID: V2178198  
02/09/2021 12:05 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.1 F 0.0  
2:00 121.5 F 0.6  
4:00 143.6 F 3.0  
6:00 177.4 F 6.4  
8:00 213.1 F 11.5  
10:00 230.8 F 7.2  
12:00 234.2 F 13.1  
14:00 236.2 F 9.8  
16:00 234.3 F 7.8  
18:00 246.9 F 14.3  
20:00 259.6 F 21.3  
22:00 271.4 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.9  
1:00 273.1 F 29.4  
2:00 271.3 F 29.1  
3:00 273.1 F 29.4  
4:00 271.1 F 29.0

Min 271.0 F 28.8  
Max 273.5 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 248.4 F 0.0  
10:00 237.9 F 0.0  
15:00 181.4 F 0.0  
20:00 173.9 F 0.0  
25:00 170.2 F 0.0  
30:00 167.4 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 22:38  
STERILIZING: 4:00  
VENTING: 2:57  
DRYING: 30:00  
TOTAL CYCLE: 01:00:31  
CYCLE COMPLETE

02/09/2021 01:05 PM  
ACCEPTED: X  
REJECTED:  
Operator: U01