
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007662
10:17 11/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:59
PRESSURIZING 2:02
STERILIZING 3:06
136.0 C 220kPa 3:06
136.1 C 222kPa 3:36
136.4 C 221kPa 4:06
136.4 C 221kPa 4:36
136.4 C 223kPa 5:06
136.4 C 222kPa 5:36
136.4 C 225kPa 6:06
136.4 C 225kPa 6:36
136.4 C 225kPa 6:36
VENTING 6:36
AIR DRYING 7:29
CYCLE COMPLETE 37:29

Digital Signature #
47EDF75C31C85893
