

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3851

Name: RITTER2  
Sterilizer ID: V2313228  
08/13/2024 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.7 F 0.0  
2:00 71.0 F 0.0  
4:00 78.4 F 0.0  
6:00 107.8 F 1.7  
8:00 142.9 F 4.4  
10:00 168.0 F 7.4  
12:00 190.2 F 10.5  
14:00 219.2 F 5.1  
16:00 217.8 F 6.6  
18:00 228.5 F 10.4  
20:00 232.2 F 8.8  
22:00 239.6 F 10.9  
24:00 234.6 F 8.2  
26:00 247.9 F 14.4  
28:00 258.4 F 20.3  
30:00 268.3 F 26.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 28.8  
1:00 271.7 F 29.0  
2:00 272.2 F 29.6  
3:00 272.1 F 29.2  
4:00 272.5 F 29.7

Min 270.9 F 28.6  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 233.1 F 0.0  
10:00 225.0 F 0.0  
15:00 181.8 F 0.0  
20:00 177.3 F 0.0  
25:00 174.3 F 0.0  
30:00 172.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:15  
STERILIZING: 4:00  
VENTING: 3:21  
DRYING: 30:00  
TOTAL CYCLE: 01:09:29  
CYCLE COMPLETE

08/13/2024 10:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_