

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4013

Name: RITTER 1  
Sterilizer ID: V2178198  
11/13/2023 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.0 F 0.0  
2:00 132.1 F 1.7  
4:00 147.0 F 4.4  
6:00 170.8 F 7.7  
8:00 197.3 F 11.8  
10:00 215.0 F 4.4  
12:00 220.8 F 8.8  
14:00 232.5 F 13.3  
16:00 235.7 F 9.5  
18:00 238.4 F 10.1  
20:00 245.5 F 12.9  
22:00 256.5 F 18.8  
24:00 266.9 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.4 F 29.2  
2:00 271.9 F 28.8  
3:00 272.7 F 29.3  
4:00 271.7 F 28.7

Min 270.8 F 28.1  
Max 273.0 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
5:00 229.0 F 0.0  
10:00 233.4 F 0.0  
15:00 226.9 F 0.0  
20:00 230.1 F 0.0  
25:00 229.6 F 0.0  
30:00 226.2 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 25:38  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:04:28  
CYCLE COMPLETE

11/13/2023 03:23 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_