
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007368
9:47 13/05/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:58
STERILIZING 4:02
136.5 C 220kPa 4:02
136.3 C 224kPa 4:32
136.2 C 223kPa 5:02
136.4 C 222kPa 5:32
136.2 C 222kPa 6:02
136.3 C 225kPa 6:32
136.4 C 222kPa 7:02
136.1 C 220kPa 7:32
136.1 C 220kPa 7:32
VENTING 7:32
AIR DRYING 8:26
CYCLE COMPLETE 8:39

Digital Signature #
8E917532ACCEA198
