
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001986
9:22 03/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:06
STERILIZING 3:11
136.0 C 220kPa 3:11
135.9 C 217kPa 3:41
136.4 C 220kPa 4:11
136.2 C 220kPa 4:41
136.3 C 222kPa 5:11
136.1 C 221kPa 5:41
136.2 C 221kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:42
AIR DRYING 7:32
CYCLE COMPLETE 8:49

Digital Signature #
79CBFA95D5A26E99
