
Midmark M11 - v1.0.0
Total Cycles: 4533

Name: RITTER 1
Sterilizer ID: V2178198
05/08/2024 08:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 73.4 F 0.0
4:00 89.4 F 0.7
6:00 128.2 F 3.0
8:00 156.4 F 5.4
10:00 176.3 F 8.7
12:00 194.4 F 12.4
14:00 215.0 F 4.0
16:00 222.1 F 7.4
18:00 231.9 F 11.2
20:00 227.6 F 6.2
22:00 240.5 F 11.5
24:00 236.4 F 8.9
26:00 249.5 F 15.0
28:00 260.7 F 21.2
30:00 271.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.5 F 29.1
2:00 272.6 F 29.2
3:00 272.1 F 29.2
4:00 271.6 F 28.8

Min 270.9 F 28.1
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 222.0 F 0.0
10:00 193.1 F 0.0
15:00 196.3 F 0.0
20:00 195.6 F 0.0
25:00 201.3 F 0.0
30:00 197.3 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:35
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:08:48
CYCLE COMPLETE

05/08/2024 09:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____