
Midmark M11 - v1.0.0
Total Cycles: 2779

Name: RITTER 1
Sterilizer ID: V2178198
01/12/2023 05:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 190.1 F 0.0
4:00 201.5 F 4.8
6:00 215.4 F 9.9
8:00 226.1 F 7.1
10:00 236.3 F 11.1
12:00 234.9 F 8.7
14:00 240.1 F 10.0
16:00 248.0 F 14.2
18:00 260.4 F 21.1
20:00 271.6 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 272.0 F 28.4
3:00 271.8 F 28.7
4:00 272.7 F 29.1

Min 271.5 F 28.2
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 222.5 F 0.0
10:00 240.0 F 0.0
15:00 234.5 F 0.0
20:00 223.2 F 0.0
25:00 223.5 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 20:39
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

01/12/2023 06:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____